


Verification Panel for Gram-Positive Bacteria Blood Culture Assays

350850
2025-07

For Research Use Only.
Not for use in diagnostic procedures.

 2026-01-07

The product may be used until expiration date.

Expected Results

Gram-Positive Bacteria			
Target	LOT 52241187	Positive Sample	LOT 52241184 Negative Sample
<i>Staphylococcus aureus</i>		Detected	Not Detected
<i>Staphylococcus epidermidis</i>		Detected	Not Detected
<i>Staphylococcus lugdunensis</i>		Detected	Not Detected
<i>Streptococcus agalactiae</i>		Detected	Not Detected
<i>Streptococcus pneumoniae</i>		Detected	Not Detected
<i>Streptococcus pyogenes</i>		Detected	Not Detected
<i>Enterococcus faecalis</i>		Detected	Not Detected
<i>Enterococcus faecium</i>		Detected	Not Detected
<i>Staphylococcus</i> spp.		Detected	Not Detected
<i>Streptococcus</i> spp.		Detected	Not Detected
<i>Listeria</i> spp.		Detected	Not Detected
<i>Streptococcus anginosus</i> group		Detected	Not Detected
<i>Bacillus</i> spp.		Detected	Not Detected

Antimicrobial Resistance Genes			
Gene	LOT 52241187	Positive Sample	LOT 52241184 Negative Sample
<i>mecA/mecC</i>		Detected	Not Reviewed
<i>vanA</i>		Detected	Not Reviewed
<i>vanB</i>		Detected	Not Reviewed

All product names, logos, brands, and marks are property of their respective holders.