


Quality Control for Gram-Positive Bacteria Blood Culture Assays

350842
2025-07

For Research Use Only.
Not for use in diagnostic procedures.

 2026-03-04

The product may be used until expiration date.

Expected Results

Gram-Positive Bacteria						
Target	LOT	52791177	Positive Control	LOT	52791173	Negative Control
<i>Staphylococcus aureus</i>			Detected			Not Detected
<i>Staphylococcus epidermidis</i>			Detected			Not Detected
<i>Staphylococcus lugdunensis</i>			Detected			Not Detected
<i>Streptococcus agalactiae</i>			Detected			Not Detected
<i>Streptococcus pneumoniae</i>			Detected			Not Detected
<i>Streptococcus pyogenes</i>			Detected			Not Detected
<i>Enterococcus faecalis</i>			Detected			Not Detected
<i>Enterococcus faecium</i>			Detected			Not Detected
<i>Staphylococcus</i> spp.			Detected			Not Detected
<i>Streptococcus</i> spp.			Detected			Not Detected
<i>Listeria</i> spp.			Detected			Not Detected
<i>Streptococcus anginosus</i> group			Detected			Not Detected
<i>Bacillus</i> spp.			Detected			Not Detected

Antimicrobial Resistance Genes						
Gene	LOT	52791177	Positive Control	LOT	52791173	Negative Control
<i>mecA/mecC</i>			Detected			Not Reviewed
<i>vanA</i>			Detected			Not Reviewed
<i>vanB</i>			Detected			Not Reviewed

All product names, logos, brands, and marks are property of their respective holders.