

Quality Control for the Luminex VERIGENE® Gram-Negative Blood Culture Test (BC-GN)<sup>1</sup>

350786-1  
2023-09

2026-07-28

Rx Only

The product may be used until expiration date<sup>2</sup>.

## Luminex VERIGENE® System

Gram-Negative Bacteria <sup>3</sup>		
Target <sup>4</sup>	[LOT] 52091159 Positive Control <sup>5</sup>	[LOT] 52091156 Negative Control <sup>6</sup>
<i>Escherichia coli</i>	Detected <sup>7</sup>	Not Detected <sup>8</sup>
<i>Klebsiella pneumoniae</i>	Detected <sup>7</sup>	Not Detected <sup>8</sup>
<i>Klebsiella oxytoca</i>	Detected <sup>7</sup>	Not Detected <sup>8</sup>
<i>Pseudomonas aeruginosa</i>	Detected <sup>7</sup>	Not Detected <sup>8</sup>
<i>Acinetobacter</i> spp.	Detected <sup>7</sup>	Not Detected <sup>8</sup>
<i>Citrobacter</i> spp.	Detected <sup>7</sup>	Not Detected <sup>8</sup>
<i>Enterobacter</i> spp.	Detected <sup>7</sup>	Not Detected <sup>8</sup>
<i>Proteus</i> spp.	Detected <sup>7</sup>	Not Detected <sup>8</sup>

Antimicrobial Resistance Genes <sup>9</sup>		
Gene <sup>10</sup>	[LOT] 52091159 Positive Control <sup>5</sup>	[LOT] 52091156 Negative Control <sup>6</sup>
CTX-M	Detected <sup>7</sup>	Not Detected <sup>8</sup>
IMP	Detected <sup>7</sup>	Not Detected <sup>8</sup>
KPC	Detected <sup>7</sup>	Not Detected <sup>8</sup>
NDM	Detected <sup>7</sup>	Not Detected <sup>8</sup>
OXA	Detected <sup>7</sup>	Not Detected <sup>8</sup>
VIM	Detected <sup>7</sup>	Not Detected <sup>8</sup>

### 1 Quality Control for the Luminex VERIGENE® BC-GN Test

Contrôle de la qualité du Luminex VERIGENE BC-GN Test / Qualitätskontrolle für das Luminex VERIGENE BC-GN Test / Controllo di qualità per il Luminex VERIGENE BC-GN Test / Control de calidad de Luminex VERIGENE BC-GN Test / Kvalitetskontroll för Luminex VERIGENE BC-GN Test

### 2 The product may be used until expiration date.

Une fois ouvert, le produit peut être utilisé jusqu'à la date de date de péremption. / Einmal geöffnet kann das Produkt bis zum angegebenen Ablaufdatum verwendet werden. / Una volta aperto, il prodotto può essere utilizzato fino alla data di scadenza. / Una vez abierto, puede utilizarse el producto hasta la fecha de caducidad. / När produkten har öppnats kan den användas fram till utgångsdatumet.

### 3 Gram-Negative Bacteria

Bactérie Gram-négative/ Gramnegative Bakterien/ Batteri Gram-negativi/ Bacterias Gramnegativas/ Gramnegativa bakterier

### 4 Target

Cible/ Ziel / Target/ Objetivo/ Mål

### 5 Positive Control

Contrôle positif/ Positive Kontrolle/ Controllo positivo/ Control Positivo/ Positiv kontroll

### 6 Negative Control

Contrôle négatif/ Negative Kontrolle/ Controllo negativo/ Control Negativo / Negativ kontroll

### 7 Detected

Défecté / Entdeckt / Rilevato / Detectado / Detekterad

### 8 Not Detected

Non détecté / Nicht entdeckt / Non rilevato / No detectado / Ej detekterad

### 9 Antimicrobial Resistance Genes

Gènes de résistance aux antimicrobiens / Antimikrobielle Resistenzgene / Geni di resistenza antimicrobica / Genes con resistencia antimicrobiana / Antimikrobiella resistensgener

### 10 Gene

Gène/ Gene/ Gene/ Gen/ Gen

### All product names, logos, brands, and marks are property of their respective holders.

Tous les noms, logos, marques et labels de produits sont la propriété de leurs propriétaires respectifs. / Alle Produktnamen, Logos, Marken und Zeichen sind Eigentum ihrer jeweiligen Besitzer. / Tutti i nomi dei prodotti, i loghi, i marchi e le marche sono di proprietà dei rispettivi titolari. / Todos los nombres de productos, logotipos, marcas comerciales y otras marcas son propiedad de sus respectivos propietarios. / Alla produktnamn, logotyper, varumärken och märken tillhör respektive innehavare.