



Quality Control for the Luminex VERIGENE[®] Gram-Negative Blood Culture Test (BC-GN)¹

350786-2
2025-10

streck.com/IFU

Rx Only

2027-01-12

The product may be used until expiration date².

Luminex VERIGENE[®] System

Gram-Negative Bacteria ³		
Target ⁴	LOT 60121159 Positive Control ⁵	LOT 60121156 Negative Control ⁶
<i>Escherichia coli</i>	Detected ⁷	Not Detected ⁸
<i>Klebsiella pneumoniae</i>	Detected ⁷	Not Detected ⁸
<i>Klebsiella oxytoca</i>	Detected ⁷	Not Detected ⁸
<i>Pseudomonas aeruginosa</i>	Detected ⁷	Not Detected ⁸
<i>Acinetobacter</i> spp.	Detected ⁷	Not Detected ⁸
<i>Citrobacter</i> spp.	Detected ⁷	Not Detected ⁸
<i>Enterobacter</i> spp.	Detected ⁷	Not Detected ⁸
<i>Proteus</i> spp.	Detected ⁷	Not Detected ⁸

Antimicrobial Resistance Genes ⁹		
Gene ¹⁰	LOT 60121159 Positive Control ⁵	LOT 60121156 Negative Control ⁶
CTX-M	Detected ⁷	Not Detected ⁸
IMP	Detected ⁷	Not Detected ⁸
KPC	Detected ⁷	Not Detected ⁸
NDM	Detected ⁷	Not Detected ⁸
OXA	Detected ⁷	Not Detected ⁸
VIM	Detected ⁷	Not Detected ⁸

1 Quality Control for the Luminex VERIGENE[®] BC-GN Test

Contrôle de la qualité du Luminex VERIGENE BC-GN Test / Qualitätskontrolle für das Luminex VERIGENE BC-GN Test / Controllo di qualità per il Luminex VERIGENE BC-GN Test / Control de calidad de Luminex VERIGENE BC-GN Test / Kvalitetskontroll för Luminex VERIGENE BC-GN Test

2 The product may be used until expiration date.

Une fois ouvert, le produit peut être utilisé jusqu'à la date de date de péremption. / Einmal geöffnet kann das Produkt bis zum angegebenen Ablaufdatum verwendet werden. / Una volta aperto, il prodotto può essere utilizzato fino alla data di scadenza. / Una vez abierto, puede utilizarse el producto hasta la fecha de caducidad. / När produkten har öppnats kan den användas fram till utgångsdatumet.

3 Gram-Negative Bacteria

Bactérie Gram-négative/ Gramnegative Bakterien/ Batteri Gram-negativi/ Bacterias Gramnegativas/ Gramnegativa bakterier

4 Target

Cible/ Ziel / Target/ Objetivo/ Mål

5 Positive Control

Contrôle positif/ Positive Kontrolle/ Controllo positivo/ Control Positivo/ Positiv kontroll

6 Negative Control

Contrôle négatif/ Negative Kontrolle/ Controllo negativo/ Control Negativo / Negativ kontroll

7 Detected

DéTECTÉ / Entdeckt / Rilevato / Detectado / Detekterad

8 Not Detected

Non détecté / Nicht entdeckt / Non rilevato / No detectado / Ej detekterad

9 Antimicrobial Resistance Genes

Gènes de résistance aux antimicrobiens / Antimikrobielle Resistenzgene / Geni di resistenza antimicrobica / Genes con resistencia antimicrobiana / Antimikrobiella resistensgener

10 Gene

Gène/ Gene/ Gene/ Gen/ Gen

All product names, logos, brands, and marks are property of their respective holders.

Tous les noms, logos, marques et labels de produits sont la propriété de leurs propriétaires respectifs. / Alle Produktnamen, Logos, Marken und Zeichen sind Eigentum ihrer jeweiligen Besitzer. / Tutti i nomi dei prodotti, i loghi, i marchi e le marche sono di proprietà dei rispettivi titolari. / Todos los nombres de productos, logotipos, marcas comerciales y otras marcas son propiedad de sus respectivos propietarios. / Alla produktnamn, logotyper, varumärken och märken tillhör respektive innehavare.