

MDx-Chex[®] for BC-GP



Quality Control for the Luminex VERIGENE[®] Gram-Positive
Blood Culture Test (BC-GP)¹

350785-3
2024-08

2027-01-12

Rx Only

The product may be used until expiration date².

Luminex VERIGENE[®] System

Gram-Positive Bacteria ³		
Target ⁴	60121155 Positive Control ⁵	60121156 Negative Control ⁶
<i>Staphylococcus aureus</i>	Detected ⁷	Not Detected ⁸
<i>Staphylococcus epidermidis</i>	Detected ⁷	Not Detected ⁸
<i>Staphylococcus lugdunensis</i>	Detected ⁷	Not Detected ⁸
<i>Streptococcus agalactiae</i>	Detected ⁷	Not Detected ⁸
<i>Streptococcus pneumoniae</i>	Detected ⁷	Not Detected ⁸
<i>Streptococcus pyogenes</i>	Detected ⁷	Not Detected ⁸
<i>Enterococcus faecalis</i>	Detected ⁷	Not Detected ⁸
<i>Enterococcus faecium</i>	Detected ⁷	Not Detected ⁸
<i>Staphylococcus ssp.</i>	Detected ⁷	Not Detected ⁸
<i>Streptococcus spp.</i>	Detected ⁷	Not Detected ⁸
<i>Listeria spp.</i>	Detected ⁷	Not Detected ⁸
<i>Streptococcus anginosus</i>	Detected ⁷	Not Detected ⁸

Antimicrobial Resistance Genes ⁹		
Gene ¹⁰	60121155 Positive Control ⁵	60121156 Negative Control ⁶
<i>mecA</i>	Detected ⁷	N/A
<i>vanA</i>	Detected ⁷	N/A
<i>vanB</i>	Detected ⁷	N/A

1 Quality Control for the Luminex VERIGENE[®] BC-GP Test

Contrôle de la qualité du Luminex VERIGENE BC-GP Test / Qualitätskontrolle für das Luminex VERIGENE BC-GP Test / Controllo di qualità per il Luminex VERIGENE BC-GP Test / Control de calidad de Luminex VERIGENE BC-GP Test / Kvalitetskontroll för Luminex VERIGENE BC-GP Test

2 The product may be used until expiration date.

Une fois ouvert, le produit peut être utilisé jusqu'à la date de date de péremption. / Einmal geöffnet kann das Produkt bis zum angegebenen Ablaufdatum verwendet werden. / Una volta aperto, il prodotto può essere utilizzato fino alla data di scadenza. / Una vez abierto, puede utilizarse el producto hasta la fecha de caducidad. / När produkten har öppnats kan den användas fram till utgångsdatumet.

3 Gram-Positive Bacteria

Bactéries Gram-positives / Grampositive Bakterien / Batteri Gram-positivi / Bacterias grampositivas / Grampositiva bakterier

4 Target

Cible/ Ziel / Target/ Objetivo/ Mål

5 Positive Control

Contrôle positif/ Positive Kontrolle/ Controllo positivo/ Control Positivo/ Positiv kontroll

6. Negative Control

Contrôle négatif/ Negative Kontrolle/ Controllo negativo/ Control Negativo / Negativ kontroll

7 Detected

Déecté / Entdeckt / Rilevato / Detectado / Detekterad

8 Not Detected

Non détecté / Nicht entdeckt / Non rilevato / No detectado / Ej detekterad

9 Antimicrobial Resistance Genes

Gènes de résistance aux antimicrobiens / Antimikrobielle Resistenzgene / Geni di resistenza antimicrobica / Genes con resistencia antimicrobiana / Antimikrobiella resistensgener

10 Gene

Gène/ Gene/ Gene/ Gen/ Gen

All product names, logos, brands, and marks are property of their respective holders.

Tous les noms, logos, marques et labels de produits sont la propriété de leurs propriétaires respectifs. / Alle Produktnamen, Logos, Marken und Zeichen sind Eigentum ihrer jeweiligen Besitzer. / Tutti i nomi dei prodotti, i loghi, i marchi e le marche sono di proprietà dei rispettivi titolari. / Todos los nombres de productos, logotipos, marcas comerciales y otras marcas son propiedad de sus respectivos propietarios. / Alla produktnamn, logotyper, varumärken och märken tillhör respektive innehavare.