

TITLE Cell-Free DNA BCT® RUO Certificate of Analysis

CERTIFICATE OF ANALYSIS

The following released product satisfies Streck product specifications for product performance and appearance.

PRODUCT NAME: Cell-Free DNA BCT® RUOLOT #: 52300315DATE OF MANUFACTURE: 2025-08-18EXPIRATION DATE: 2027-08-18ACTIVITY*(Specification Range 2.30 – 3.46): 2.99QUALITY CONTROL APPROVAL: Cynthia Wong

(Print)

QUALITY CONTROL APPROVAL: 

(Signature)

TITLE: Quality Control Product SpecialistDATE: 2025-12-11

This product was manufactured in compliance with all current, applicable regulatory requirements and standards. The product is For Research Use Only. Not for use in diagnostic procedures.

This product has been validated per the ISO 6710 standard.

This product has been sterilized. The Sterilization Process has been established and validated in accordance with ISO 11137, Sterilization of Healthcare Products. All established dosage requirements are traceable to this internationally recognized standard.

*Reflects the active ingredient concentration in Cell-Free DNA BCT as measured via Streck's proprietary test method.