




Quality Control for the Luminex VERIGENE<sup>®</sup> Gram-Positive  
Blood Culture Test (BC-GP)<sup>1</sup>

350785-3  
2024-08

 2026-02-10

Rx Only

The product may be used until expiration date<sup>2</sup>.

Luminex VERIGENE<sup>®</sup> System

Gram-Positive Bacteria <sup>3</sup>		
Target <sup>4</sup>	LOT 50411155 Positive Control <sup>5</sup>	LOT 50411156 Negative Control <sup>6</sup>
<i>Staphylococcus aureus</i>	Detected <sup>7</sup>	Not Detected <sup>8</sup>
<i>Staphylococcus epidermidis</i>	Detected <sup>7</sup>	Not Detected <sup>8</sup>
<i>Staphylococcus lugdunensis</i>	Detected <sup>7</sup>	Not Detected <sup>8</sup>
<i>Streptococcus agalactiae</i>	Detected <sup>7</sup>	Not Detected <sup>8</sup>
<i>Streptococcus pneumoniae</i>	Detected <sup>7</sup>	Not Detected <sup>8</sup>
<i>Streptococcus pyogenes</i>	Detected <sup>7</sup>	Not Detected <sup>8</sup>
<i>Enterococcus faecalis</i>	Detected <sup>7</sup>	Not Detected <sup>8</sup>
<i>Enterococcus faecium</i>	Detected <sup>7</sup>	Not Detected <sup>8</sup>
<i>Staphylococcus ssp.</i>	Detected <sup>7</sup>	Not Detected <sup>8</sup>
<i>Streptococcus spp.</i>	Detected <sup>7</sup>	Not Detected <sup>8</sup>
<i>Listeria spp.</i>	Detected <sup>7</sup>	Not Detected <sup>8</sup>
<i>Streptococcus anginosus</i>	Detected <sup>7</sup>	Not Detected <sup>8</sup>

Antimicrobial Resistance Genes <sup>9</sup>		
Gene <sup>10</sup>	LOT 50411155 Positive Control <sup>5</sup>	LOT 50411156 Negative Control <sup>6</sup>
<i>mecA</i>	Detected <sup>7</sup>	N/A
<i>vanA</i>	Detected <sup>7</sup>	N/A
<i>vanB</i>	Detected <sup>7</sup>	N/A

## 1 Quality Control for the Luminex VERIGENE<sup>®</sup> BC-GP Test

Contrôle de la qualité du Luminex VERIGENE BC-GP Test / Qualitätskontrolle für das Luminex VERIGENE BC-GP Test / Controllo di qualità per il Luminex VERIGENE BC-GP Test / Control de calidad de Luminex VERIGENE BC-GP Test / Kvalitetskontroll för Luminex VERIGENE BC-GP Test

## 2 The product may be used until expiration date.

Une fois ouvert, le produit peut être utilisé jusqu'à la date de date de péremption. / Einmal geöffnet kann das Produkt bis zum angegebenen Ablaufdatum verwendet werden. / Una volta aperto, il prodotto può essere utilizzato fino alla data di scadenza. / Una vez abierto, puede utilizarse el producto hasta la fecha de caducidad. / När produkten har öppnats kan den användas fram till utgångsdatumet.

## 3 Gram-Positive Bacteria

Bactéries Gram-positives / Grampositive Bakterien / Batteri Gram-positivi / Bacterias grampositivas / Grampositiva bakterier

## 4 Target

Cible/ Ziel / Target/ Objetivo/ Mål

## 5 Positive Control

Contrôle positif/ Positive Kontrolle/ Controllo positivo/ Control Positivo/ Positiv kontroll

## 6. Negative Control

Contrôle négatif/ Negative Kontrolle/ Controllo negativo/ Control Negativo / Negativ kontroll

## 7 Detected

Défecté / Entdeckt / Rilevato / Detectado / Detekterad

## 8 Not Detected

Non détecté / Nicht entdeckt / Non rilevato / No detectado / Ej detekterad

## 9 Antimicrobial Resistance Genes

Gènes de résistance aux antimicrobiens / Antimikrobielle Resistenzgene / Geni di resistenza antimicrobica / Genes con resistencia antimicrobiana / Antimikrobiella resistensgener

## 10 Gene

Gène/ Gene/ Gen/ Gen/ Gen

## All product names, logos, brands, and marks are property of their respective holders.

Tous les noms, logos, marques et labels de produits sont la propriété de leurs propriétaires respectifs. / Alle Produktnamen, Logos, Marken und Zeichen sind Eigentum ihrer jeweiligen Besitzer. / Tutti i nomi dei prodotti, i loghi, i marchi e le marche sono di proprietà dei rispettivi titolari. / Todos los nombres de productos, logotipos, marcas comerciales y otras marcas son propiedad de sus respectivos propietarios. / Alla produktnamn, logotyper, varumärken och märken tillhör respektive innehavare.