

Whole Blood Assayed Hematology Control¹

Rx Only

Open-vial stability 95 days²



2025-06-24

350206-40
2023-09

Automated Methods³

	[LOT] L1 - 41761380		[LOT] L2 - 41761381	
	\bar{x}	+/-	\bar{x}	+/-
Instrument ⁴	mm / hr	mm / hr	mm / hr	mm / hr
Vital Diagnostics Excyte 10 / M	9	8	70	18

Manual Methods⁵

	[LOT] L1 - 41761380		[LOT] L2 - 41761381	
	\bar{x}	+/-	\bar{x}	+/-
Instrument ⁴	mm / hr	mm / hr	mm / hr	mm / hr
Classical Westergren ^a	5	4	15	7
Modified Westergren - Globe Sedi-Rate	5	4	29	20
Modified Westergren - Polymedco Sediplast	5	4	39	20
Modified Westergren - Streck ESR-10 Manual Rack	9	8	55	18
Wintrobe	5	4	16	7

a Undiluted Westergren Method

Valeurs principales assignées / Unverdünnte Westergren-Methode / Metodo Westergren non diluito / Método Westergren sin diluir / Outspädd Westergrenmetod

1 Whole Blood Assayed Hematology Control

Contrôle d'hématologie dosé « sang total » / Mit Vollblut getestete Hämatologiekontrolle / Controllo ematologico per analisi di sangue intero / Control de hematologia de sangre entera analizada / Helblodsanalyserad hematologikontroll

2 Open-vial stability 95 days

Stabilité en flacon ouvert 95 jours / Stabilität geöffneter Flaschen 95 tage / Stabilità della fiala aperta 95 giorni / Estabilidad de la cápsula abierta 95 días / Hållbarhet för öppen flaska 95 dagar

3 Automated Methods

Méthodes automatisées / Automatische Verfahren / Metodi automatizzati / Métodos automatizados / Automatiska metoder

4 Instrument

Instrument / Gerät / Strumento / Instrumento / Instrument

5 Manual Methods

Méthodes manuelles / Manuelle Verfahren / Metodi manuali / Métodos manuales / Manuella metoder

\bar{x} Mean

Moyenne / Mittelwert / Media / Media / Medelvärde

+/- Expected Range

± Intervalle escompté / ± Erwartungsbereich / ± Range previsto / ± Intervalo previsto / ± Förväntat intervall