

TITLE

CD-Chex Plus® Certificate of Analysis

Date of Manufacture: 2026-06-15

Expiration Date: 2026-09-26

Labeled as Lot #:

61800072

61800071

Hematological Analysis

WBC (x10⁹/L):

6.4

6.3

Flow Cytometry Analysis:

Pass

Pass

Debris:

Pass

Pass

Staining Results:

Pass

Pass

Microscopic Examination:

Pass

Pass

Values obtained from Sysmex® XP-300, BD FACSCanto™ II

The finished product does not contain any actively growing yeasts, mold, gram negative or gram positive microorganisms as determined by standard aerobic plate count.

CAUTION: All blood products should be treated as potentially infectious. Source material from which this product was derived was found negative when tested in accordance with current FDA required tests. No known test methods can offer assurance that products derived from human blood will not transmit infectious agents. See the Instructions (IFU) tab under Resources on the product page at streck.com for specific FDA required blood tests.

This product was manufactured in compliance with all current, applicable regulatory requirements and standards.

To the best knowledge of the Streck Quality Control Department, the information above is true and accurate.

Streck Quality Control Department

Signature: 

Date: 2026-06-15