

Quality Control and Verification Panel for
Diasorin LIAISON PLEX[®] Gram-Negative Blood Culture Assay

350833
2026-05

 streck.com/IFU

Rx Only

 2027-06-15

The product may be used until expiration date.

Expected Results

Gram-Negative Bacteria					
Target	LOT	61661160	Control 1	LOT	61661161 Control 2
<i>Enterobacteriaceae/Morganellaceae</i>		Not Detected			Detected
<i>Acintebacter</i> spp.		Detected			Not Detected
<i>Acintebacter baumannii</i>		Detected			Not Detected
<i>Citrobacter</i> spp.		Not Detected			Detected
<i>Enterobacter</i> spp.		Not Detected			Detected
<i>Escherichia coli</i>		Not Detected			Detected
<i>Haemophilus influenzae</i>		Detected			Not Detected
<i>Klebsiella oxytoca</i>		Not Detected			Detected
<i>Klebsiella pneumoniae</i>		Not Detected			Detected
<i>Klebsiella variicola</i>		Not Detected			Detected
<i>Morganella morganii</i>		Not Detected			Detected
<i>Neisseria meningitidis</i>		Detected			Not Detected
<i>Proteus</i> spp.		Not Detected			Detected
<i>Pseudomonas</i> spp.		Detected			Not Detected
<i>Pseudomonas aeruginosa</i>		Detected			Not Detected
<i>Salmonella</i> spp.		Not Detected			Detected
<i>Serratia marcescens</i>		Not Detected			Detected
<i>Stenotrophomonas maltophilia</i>		Detected			Not Detected
Antimicrobial Resistance Genes					
Gene	LOT	61661160	Control 1	LOT	61661161 Control 2
CTX-M		Not Detected			Detected
IMP		Not Detected			Detected
KPC		Detected			Not Detected
MCR		Not Detected			Detected
NDM		Detected			Not Detected
OXA		Not Detected			Detected
SME		Not Reviewed			Detected
VIM		Detected			Not Detected

All product names, logos, brands, and marks are property of their respective holders.