
TITLE CERTIFICATE OF ANALYSIS – CELL-FREE DNA BCT CE REAGENT

CERTIFICATE OF ANALYSIS

The following released product satisfies Streck product specifications for product performance and appearance.

PRODUCT NAME: Cell-Free DNA BCT[®] CE Reagent

LOT #: 60830317

DATE OF MANUFACTURE: 2016-02-29

EXPIRATION DATE: 2018-03-23

ACTIVITY*(Specification Range 2.30 – 3.46): 2.85

This product was manufactured under FDA QSR, ISO 9001/ISO 13485 and CMDR conditions.

This product is FOR EXPORT ONLY, not to be sold in the United States.

*Reflects the active ingredient concentration in Cell-Free DNA BCT CE as measured via Streck's proprietary test method.